

# PTRANS-T10

Single-Use  
Temperature Logger

## 1. Start Recording

Keep pressing the START button for 3 seconds to start the recording, the green LED will turn on for 1.5s after the logger have been successful startup.

After the product is started, it will enter the start-up delay state, and the green light will light up every 5 seconds.

When the logger formally enters the recording state, according to the monitoring situation, if the over temperature alarm, the red LED will flash once in 10 seconds, if the temperature is normal, the green LED will flash once in 10 seconds.



## 2. Stop Recording

You can stop the logger by insert into the computer USB interface.

When there is no operation thereafter, the product will indicate the recorded result every 10 seconds. If the recorded result is Pass, the green LED will flash, otherwise the red LED will flash.

## 3. Download The Report

When you need to download the report after the trip, plug the logger into the USB port of your computer, the green lights of the product flash during the report generation process, and the red indicator light will be on when the report is generated successfully, so that you can read the report at the "POLWAX" removable disk icon which is displayed on the computer desktop.

Note: After generating a report, the logger cannot record data again and can only read the report stored in it.

## 4. Specification Parameters

LED Indicator Green (Recording, OK) , Red (Alarm)

Operation Temperature  $-30^{\circ}\text{C} \sim 70^{\circ}\text{C}$

Measurement  $\pm 0.5^{\circ}\text{C}$  @  $-10 \sim 40^{\circ}\text{C}$ ,  
Accuracy  $\pm 0.8^{\circ}\text{C}$  @  $-30 \sim -10^{\circ}\text{C}$  &  $40 \sim 70^{\circ}\text{C}$

Resolution  $0.1^{\circ}\text{C}$

Memory Space Up to 33,000 points

Start Mode Manual start

IP Grade IP67

Battery Type 3.0V CR2032

Weight 20g

Connection USB2.0

Dimensions 70\*27\*7mm

